

Application Sheet

HSV 2 IgG

REF 51226

Instrument: Elisys Duo

REF 17200

Setting:

HSV2G_51226_R4.asy

For all essential product information, please refer to the User Manual of the analyser and the Instructions for Use for reagents, controls and calibrators. Please refer to the Instructions for Use for performance data.

This Application Sheet provides additional information regarding the use of the assay on the HUMAN ELISA analyser.

The parameters defined in this application sheet have been developed to provide optimal product performance with the assay and instrument combination. Any modification to these parameters may negatively affect the performance and results of the assay. The user is responsible for the validation of any modification to the protocol here described.

Material Required

Material	REF	Comments
HSV 2 IgG	51226	
Sample Rack Elisys Duo	17200/23	NC, PC, sample collection tubes
Reagent Rack (2 Lanes) Elisys Duo	17200/25	CON, SUB, STOP
Adaptor 100 ml Square Bottle	17200/30	DIL
Dilution Plate	17200/255	predilution of samples

Additional Notes

OD-values of 3.5 or above are outside the photometer's dynamic range. For this reason, they are displayed as 4.0.

When processing a whole plate, it is recommended to use completely filled tip racks. To refill incomplete racks in the software, press the button "Refill partial tip racks" in the Load screen before starting the run. Otherwise the tip racks will have to be refilled several times during the run, which will result in avoidable interruptions.