

# Application Sheet

## IMTEC-MPO-ANCA

**REF** ITC82030

## Instrument: Elisys Quattro

**REF** 16300

### Setting:

**MPO\_ITC82030\_R4.asy**

For all essential product information, please refer to the User Manual of the analyser and the Instructions for Use for reagents, controls and calibrators. Please refer to the Instructions for Use for performance data.

This Application Sheet provides additional information regarding the use of the assay on the HUMAN ELISA analyser.

The parameters defined in this application sheet have been developed to provide optimal product performance with the assay and instrument combination. Any modification to these parameters may negatively affect the performance and results of the assay. The user is responsible for the validation of any modification to the protocol here described.

### Material Required

Material	REF	Comments
IMTEC-MPO-ANCA	ITC82030	
Reagent Rack 0 Elisys Quattro	16300/38	All reagents, except DIL
Reagent Rack 2	16300/40	Suitable for sample collection tubes
Sample Rack 20 Elisys Quattro	16300/41	Predilution of samples

### Additional Notes

The concentration of the PC must be checked manually since the target values are batch specific. The concentration of the NC must also be checked by the user.

OD-values of 3.5 or above are outside the photometer's dynamic range. For this reason, they are displayed as 4.0.