

Application Sheet

fPSA

REF 52035

Instrument: Elisis Quattro

REF 16300

Setting:

fPSA_52035_R5.asy

For all essential product information, please refer to the User Manual of the analyzer and the instructions for use for reagents, controls, calibrators and the Certificate of Analysis. Please refer to the instructions for use for performance data.

This Application Sheet provides additional information regarding the use of the assay on the HUMAN ELISA analyzer.

The parameters defined in this application sheet have been developed to provide optimal product performance with the assay and instrument combination. Any modification to these parameters may negatively affect the performance and results of the assay. The user is responsible for the validation of any modification to the protocol here described.

Material Required

Material	REF	Comments
fPSA	52035	
Reagent Rack II Elisis Quattro	16300/39	Cal A – Cal F, CON; SUB; STOP
Sample Rack 20 Elisis Quattro	16300/41	Suitable for sample collection tubes

Additional Notes

The substrate is filled to the bottleneck. Due to different diameters of neck and body of the bottle the fluid level decreases faster initially, which can lead to the suction of air bubbles. Therefore, before starting the first measurement using a fresh kit, 1 ml of substrate has to be removed from the substrate bottle and should be discarded.