

# Application Sheet

## anti-HBs

**REF 51255**

## Instrument: Elisys Uno

**REF 17350**

### Setting:

**A-HBS\_51255\_R4**

For all essential product information, please refer to the User Manual of the analyzer and the instructions for use for reagents, controls, calibrators and the Certificate of Analysis. Please refer to the instruction for use for performance data.

This Application Sheet provides additional information regarding the use of the assay on the HUMAN ELISA analyzer.

The parameters defined in this application sheet have been developed to provide optimal product performance with the assay and instrument combination. Any modification to these parameters may negatively affect the performance and results of the assay. The user is responsible for the validation of any modification to the protocol here described.

### Material Required

Material	REF	Comments
anti-HBs	51255	
Reagent Rack, Hormone Elisys Uno	17359/5	all reagents
Sample Rack (13 mm) Elisys Uno	17359/8	Sample, according to sample collection tube

### Additional Notes

A combined processing with other assays is not possible because of the required incubation conditions.

Sample dispensing is performed at room temperature. At the beginning of the incubation step the final temperature of 37°C is not reached. Hence, the instrument's message "Temperature out of Range" and the corresponding note "T\_low" do not indicate an error. It is part of the regular procedure.