

# Application Sheet

**fPSA**

**REF 52035**

**Instrument: Elisys Uno**

**REF 17350**

**Setting:**

**fPSA\_52035\_R5**

For all essential product information, please refer to the User Manual of the analyzer and the instructions for use for reagents, controls, calibrators and the Certificate of Analysis. Please refer to the instructions for use for performance data.

This Application Sheet provides additional information regarding the use of the assay on the HUMAN ELISA analyzer.

The parameters defined in this application sheet have been developed to provide optimal product performance with the assay and instrument combination. Any modification to these parameters may negatively affect the performance and results of the assay. The user is responsible for the validation of any modification to the protocol here described.

## Material Required

Material	REF	Comments
fPSA	52035	
Reagent Rack Hormone	17359/5	all reagents
Sample Rack (13 mm) Elisys Uno	17359/8	Sample, according to sample collection tube

## Additional Notes

For technical reasons the assay step „start incubation time” requires to define a corresponding reagent. In the instrument software the reagent „start incubation time“ can be allocated to any position on the reagent rack. Do not place a bottle to the position on the reagent rack.