

Application Sheet

25-OH-Vitamin D

REF 55500

Instrument: HumaReader HS

REF 16670

Setting:

VitD_90m_55500_R2

For all essential product information, please refer to the User Manual of the analyzer and the instructions for use for reagents, controls, calibrators and the Certificate of Analysis. Please refer to the instruction for use for performance data.

This Application Sheet provides additional information regarding the use of the assay on the HUMAN ELISA analyzer.

The parameters defined in this application sheet have been developed to provide optimal product performance with the assay and instrument combination. Any modification to these parameters may negatively affect the performance and results of the assay. The user is responsible for the validation of any modification to the protocol here described.

Additional Notes

Calibrator concentrations are batch specific. The user has to verify and edit the values that are deposited in the device.

The concentration of control 1 and control 2 must be checked by the user. Target values are batch specific.