

Application Sheet

HSV 1 IgG

REF 51216

Instrument: HumaReader Single plus

REF 18000

Setting:

HSV1G_51216_R4

For all essential product information, please refer to the User Manual of the analyser and the Instructions for Use for reagents, controls and calibrators. Please refer to the Instructions for Use for performance data.

This Application Sheet provides additional information regarding the use of the assay on the HUMAN ELISA analyser.

The parameters defined in this application sheet have been developed to provide optimal product performance with the assay and instrument combination. Any modification to these parameters may negatively affect the performance and results of the assay. The user is responsible for the validation of any modification to the protocol here described.

Additional Notes

The criteria for Blank, NC, PC are checked automatically. If the criteria fail because of invalid PC and NC the reading is stopped and no results will be displayed.