

Immunoassay Multi Control HumaCLIA SR

REF 84850

LOT 23001

Version 06 / 02 - 2025

Target Values / Sollwerte / Valores asignados / Valeurs cibles

 2026-12-06

If a reagent kit with a new LOT number is used, please check the link for latest version of target value sheet:



<https://www.human.de/clia/target-value-sheets/>

REF	Assay Parameter Ensayo Essai	Valid from Gültig ab Válido a partir de Valable à partir de [RGT] Lot	Units Einheiten Unidades Unité	Target Sollwert Valor asignado Cible	SD SD SD ET	Range Bereich Intervalo Intervalle	Target Sollwert Valor asignado Cible	SD SD SD ET	Range Bereich Intervalo Intervalle
				Level 1			Level 2		
84600	TSH	24001	µIU/ml	4.07	0.53	2.47 - 5.66	36.10	3.36	26.02 - 46.17
84600	TSH	24003	µIU/ml	3.15	0.41	1.91 - 4.39	27.36	2.54	19.73 - 34.99
84615	T3	24001	ng/ml	1.09	0.14	0.66 - 1.52	3.39	0.44	2.06 - 4.72
84615	T3	25001	ng/ml	1.23	0.16	0.75 - 1.71	3.32	0.44	2.01 - 4.63
84620	fT3	24001	pg/ml	3.30	0.43	2.00 - 4.60	8.91	0.83	6.42 - 11.40
84620	fT3	25001	pg/ml	3.34	0.44	2.03 - 4.65	12.30	1.14	8.87 - 15.73
84610	fT4	24001	pg/ml	10.03	1.32	6.08 - 13.98	37.20	3.46	26.82 - 47.57
84610	fT4	24002	pg/ml	12.39	1.63	7.51 - 17.27	31.67	2.95	22.83 - 40.51
82120	FSH	24001	mIU/ml	4.97	0.65	3.01 - 6.92	38.21	3.55	27.55 - 48.87
82115	LH	24001	mIU/ml	4.92	0.64	2.98 - 6.85	66.81	6.21	48.17 - 85.45
84605	T4	24001	ng/ml	66.10	6.15	47.66 - 84.54	166.14	15.45	119.78 - 212.49
84625	anti-TG	24001	IU/ml	31.81	2.96	22.94 - 40.68	102.50	9.53	73.90 - 131.10
84625	anti-TG	24002	IU/ml	30.87	2.87	22.26 - 39.48	100.57	9.35	72.51 - 128.62
84630	anti-TPO	24001	U/ml	18.88	1.76	13.61 - 24.14	27.53	2.56	19.85 - 35.20
83100	AFP	24001	ng/ml	6.42	0.84	3.90 - 8.94	117.14	10.89	84.47 - 149.81
83100	AFP	25001	ng/ml	6.49	0.85	3.94 - 9.04	118.52	11.02	85.45 - 151.59
82100	HCG	24001	mIU/ml	2.85	0.37	1.73 - 3.97	95.11	8.85	68.57 - 121.65
85000	25-OH Vitamin D	24001	ng/ml	31.36	2.92	22.61 - 40.11	46.19	4.30	33.30 - 59.07
85055	Vitamin B12	24001	pg/ml	412.80	38.39	297.63 - 527.97	1080.10	100.45	778.75 - 1381.45
82105	Progesterone	24001	ng/ml	1.16	0.15	0.69 - 1.62	28.57	2.66	20.60 - 36.53
82105	Progesterone	24002	ng/ml	1.18	0.16	0.71 - 1.65	31.87	2.96	22.98 - 40.76
83125	fPSA*	24001	ng/ml	0.77	0.10	0.47 - 1.07	3.66	0.48	2.22 - 5.10
83130	PSA*	24001	ng/ml	3.69	0.48	2.24 - 5.14	16.24	1.51	11.71 - 20.77
82125	Prolactin	24001	ng/ml	11.20	1.47	6.79 - 15.60	21.62	2.01	15.59 - 27.65
82110	Testosterone	24001	ng/ml	7.64	1.00	4.63 - 10.65	2.16	0.28	1.31 - 3.01
82110	Testosterone	24003	ng/ml	7.01	0.92	4.25 - 9.77	1.83	0.24	1.11 - 2.55
85030	Troponin I	24001	ng/ml	0.05	0.04	0.00 - 0.16	10.61	0.99	7.65 - 13.57
85040	CK-MB	24001	ng/ml	4.63	0.61	2.81 - 6.45	14.36	1.34	10.35 - 18.37
85035	Myoglobin	24001	ng/ml	33.60	3.12	24.23 - 42.97	134.77	12.53	97.17 - 172.37

REF	Assay Parameter Ensayo Essai	Valid from Gültig ab Válido a partir de Valable à partir de RGT Lot	Units Einheiten Unidades Unité	Target Sollwert Valor asignado Cible	SD SD SD ET	Range Bereich Intervalo Intervalle	Target Sollwert Valor asignado Cible	SD SD SD ET	Range Bereich Intervalo Intervalle
				Level 1			Level 2		
83110	CA 125	24001	U/ml	36.62	3.41	26.40 - 46.84	121.20	11.27	87.39 - 155.01
83115	CA 15-3	24001	U/ml	29.15	2.71	21.02 - 37.28	73.09	6.80	52.70 - 93.48
83120	CA 19-9	24001	U/ml	26.43	2.46	19.06 - 33.80	103.75	9.65	74.80 - 132.70
83120	CA 19-9	24002	U/ml	31.12	2.89	22.44 - 39.80	93.76	8.72	67.60 - 119.92
82130	Estradiol	24001	pg/ml	62.44	5.81	45.02 - 79.86	261.84	24.35	188.79 - 334.89
83105	CEA	24001	ng/ml	3.23	0.42	1.96 - 4.50	10.04	1.32	6.09 - 13.99
83105	CEA	24002	ng/ml	3.12	0.41	1.89 - 4.35	9.40	1.23	5.70 - 13.10
85010	PTH*	for 24001 only	pg/ml	31.53	2.93	22.73 - 40.33	222.68	20.71	160.55 - 284.81
85010	PTH*	24002	pg/ml	29.40	2.73	21.19 - 37.60	208.78	19.42	150.53 - 267.03
85049	Ferritin*	23002	ng/ml	141.27	13.14	101.86 - 180.68	223.86	20.82	161.40 - 286.31

* No CE – only for use outside Europe