


PCT Control HumaCLIA SR

REF 85820

LOT 23001

Version 02 / 12 - 2024

Target Values / Sollwerte / Valores asignados / Valeurs cibles

 2026-10-12

If a reagent kit with a new LOT number is used, please check the link for latest version of target value sheet:

<https://www.human.de/clia/target-value-sheets/>



REF	Assay Parameter Ensayo Essai	Valid from Gültig ab Válido a partir de Valable à partir de [RGT] Lot	Units Einheiten Unidades Unité	Target Sollwert Valor asignado Cible	SD SD SD ET	Range Bereich Intervalo Intervalle	Target Sollwert Valor asignado Cible	SD SD SD ET	Range Bereich Intervalo Intervalle
				Level 1			Level 2		
85020	PCT	23001	ng/ml	0.59	0.06	0.41 - 0.77	12.54	1.25	8.78 - 16.30
85020	PCT	24003	ng/ml	0.58	0.08	0.35 - 0.81	10.52	0.98	7.58 - 13.46