

HbA1c liquidirect Calibrator Set

LOT 24002

Version 1 / 29-04-2024

REF 10776

 30.06.2026

Link for latest version of target value sheet:

Content 4 x 0.5 ml

<https://www.human.de/target-value-sheets/>

Calibration Values / Kalibrationswerte / Valores de calibración / Valeurs de calibration

			CAL0	CAL1	CAL2	CAL3	CAL4
Device Gerät Instrumento Equipement	Reference material/method Referenzmaterial/-methode Material/método de referencia Matériel/méthode de référence	Unit Einheit Unidad Unité	Assigned Value Zugewiesener Wert Valor asignado Valeur assignée	Assigned Value Zugewiesener Wert Valor asignado Valeur assignée	Assigned Value Zugewiesener Wert Valor asignado Valeur assignée	Assigned Value Zugewiesener Wert Valor asignado Valeur assignée	Assigned Value Zugewiesener Wert Valor asignado Valeur assignée
HumaStar 100 HumaStar 200 HumaStar 300SR	IFCC HbA1c system	mmol/mol Hb	0	36.0	75.0	108	144
		%	0	5.4	9.0	12.0	15.3
HumaStar 600	IFCC HbA1c system	mmol/mol Hb	18.0	45.0	85.0	120	154
		%	3.8	6.3	9.9	13.1	16.2
other	IFCC HbA1c system	mmol/mol Hb	0	33.0	70.0	108	144
		%	0	5.2	8.6	12.0	15.3

CA-HbA1c

Rev. 007 | valid of 15.11.2022

INF 1077602 - Version 1 / 29-04-2024

Human

Human Gesellschaft für Biochemica und Diagnostica mbH
Max-Planck-Ring 21 · 65205 Wiesbaden · Germany

Telefon +49 6122-9988-0 · Telefax +49 612-9988-100 · e-Mail human@human.de