


HbA1c liquidirect Controls

LOT 23003

Version 1 /09 - 2023

REF 10775

 2025-04-30

Link for latest version of target value sheet:

<https://www.human.de/target-value-sheets/>

Target Values / Sollwerte / Valores Asignados/ Valeurs Cibles

Device	Unit
Gerät	Einheit
Instrumento	Unidad
Equipement	Unité

Control Blood normal CBN	
Mean	Range
Mittelwert	Bereich
Media	Rango
Moyenne	Plage de valeurs

Control Blood abnormal CBA	
Mean	Range
Mittelwert	Bereich
Media	Rango
Moyenne	Plage de valeurs

HumaStar 300SR	mmol/mol Hb
	%
HumaStar 600	mmol/mol Hb
	%
Other	mmol/mol Hb
	%

36.1	29.6	—	42.6
5.5	4.9	—	6.1
44.4	40.3	—	48.5
6.2	5.8	—	6.6
36.5	29.4	—	43.6
5.5	4.8	—	6.1

108	101	—	114
12.0	11.4	—	12.6
120	113	—	128
13.1	12.5	—	13.9
112	97.3	—	126
12.4	11.1	—	13.7