

Standard RHEUMATOID FACTORS

Version 1/04 - 2024

REF

11361

LOT

24003



2025 - 10 - 31

Link for latest version of target value sheet:

<https://www.human.de/target-value-sheets/>

Calibration Values / Kalibrationswerte / Valores de calibración / Valeurs de calibration

Device Gerät Instrumento Equipement	Reference material/method Referenzmaterial/-methode Material/método de referencia Matériel/méthode de référence	Unit Einheit Unidad Unité	Assigned Value Zugewiesener Wert Valor asignado Valeur assignée
RHEUMATOID FACTORS			
Multipurpose reagent, REF 11261PA			
Multipurpose reagent on HumaStar 100 & 200 LOT ≥ 23501	Adjusted to 2nd British Standard NIBSC 0064/003	IU/ml	180
		kIU/l	180
Multipurpose reagent on Humalyzer 4000 LOT ≥ 23501	Adjusted to 2nd British Standard NIBSC 0064/003	IU/ml	180
		kIU/l	180
System reagents, REF 11261300, 11261600			
HumaStar System reagent on HumaStar 300SR LOT ≥ 23501	Adjusted to 2nd British Standard NIBSC 0064/003	IU/ml	180
		kIU/l	180
HumaStar System reagent on HumaStar 600 LOT ≥ 23501	Adjusted to 2nd British Standard NIBSC 0064/003	IU/ml	180
		kIU/l	180

CA_RF_11361

Rev. 003 | valid of 04.10.2023

INF 1136102 - Version 1/04 - 2024



Human Gesellschaft für Biochemica und Diagnostica mbH
Max-Planck-Ring 21 · 65205 Wiesbaden · Germany
Telefon +49 6122-9988-0 · Telefax +49 612-9988-100 · e-Mail human@human.de