


HbA1c liquidirect Controls

LOT 24001

Version 1 /11 - 2023

REF 10775

 2026-07-31

Link for latest version of target value sheet:

<https://www.human.de/target-value-sheets/>

Target Values / Sollwerte / Valores Asignados/ Valeurs Cibles

Device Gerät Instrumento Equipement		Unit Einheit Unidad Unité	Control Blood normal CBN				Control Blood abnormal CBA			
			Mean Mittelwert Media Moyenne	Range Bereich Rango Plage de valeurs			Mean Mittelwert Media Moyenne	Range Bereich Rango Plage de valeurs		
HumaStar 300SR	mmol/mol Hb		30.1	25.6	—	34.6	109	100	—	119
	%		4.9	4.5	—	5.3	12.1	11.3	—	13.0
HumaStar 600	mmol/mol Hb		38.9	34.3	—	43.5	122	108	—	136
	%		5.7	5.3	—	6.1	13.3	12.0	—	14.6
other	mmol/mol Hb		30.9	24.2	—	37.6	115	91.9	—	138
	%		5.0	4.4	—	5.6	12.7	10.6	—	14.8

CS-HbA1c

Rev. 007 | valid of 15.11.2022

INF 1077502 - Version 1 /11 - 2023



Human Gesellschaft für Biochemica und Diagnostica mbH

Max-Planck-Ring 21 · 65205 Wiesbaden · Germany

Telefon +49 6122-9988-0 · Telefax +49 6122-9988-100 · e-Mail human@human.de