


# HbA1c liquidirect Controls

**LOT** 24002

Version 1 /05 - 2024

**REF** 10775

 2027-02-28

Link for latest version of target value sheet:

<https://www.human.de/target-value-sheets/>

Target Values / Sollwerte / Valores Asignados/ Valeurs Cibles

Device Gerät Instrumento Equipement		Unit Einheit Unidad Unité	Control Blood normal <b>CBN</b>				Control Blood abnormal <b>CBA</b>			
			Mean Mittelwert Media Moyenne	Range Bereich Rango Plage de valeurs			Mean Mittelwert Media Moyenne	Range Bereich Rango Plage de valeurs		
HumaStar 300SR	mmol/mol Hb		35.2	28.7	—	41.7	109	98.0	—	121
	%		5.4	4.8	—	6.0	12.1	11.1	—	13.2
HumaStar 600	mmol/mol Hb		43.4	36.6	—	50.3	116	105	—	127
	%		6.1	5.5	—	6.8	12.8	11.8	—	13.8
other	mmol/mol Hb		35.2	28.7	—	41.7	109	98.0	—	121
	%		5.4	4.8	—	6.0	12.1	11.1	—	13.2

CS-HbA1c

Rev. 007 | valid of 15.11.2022

INF 1077502 - Version 1 /05 - 2024



Human Gesellschaft für Biochemica und Diagnostica mbH

Max-Planck-Ring 21 · 65205 Wiesbaden · Germany

Telefon +49 6122-9988-0 · Telefax +49 6122-9988-100 · e-Mail human@human.de